

[View on rigaku.com](https://www.rigaku.com)

EDXRF1385 - Agricultural Soils & Plant Materials

Scope

The analysis of agricultural soils and plant materials is demonstrated.

Background



In the agri-food sector it is important to not only study the soil composition and use of fertilizers, but also the uptake of nutrients and potentially toxic elements within the plants and crops themselves. XRF (X-ray Fluorescence) is an accepted technique in the industry. The Rigaku NEX CG meets the challenges of soil and crop analysis using indirect excitation EDXRF (Energy Dispersive XRF). Secondary targets and polarization give the operator high precision instrumentation with a simple and intuitive software design ideal to meet the demanding trace measurements as well as the measurements of major and minor elements.



Samples

Samples were provided by a major European advisory authority for agricultural and food development and testing.

Results: Plant materials

Plant material and crops are mainly organic material with low inorganic levels exhibiting minimal or no mineralogical effects. Three samples of plant material were provided. Two of the samples were used to create a Matching Library, a procedure that is simple for the user to build. The remaining sample was measured as an unknown with results shown here.

Plant Material			
Units: ppm (mg/kg)			
Element	Reference value	NEX CG result	Stat. error
Na	752	859	121
Mg	3230	3410	89
Al	---	1250	11
Si	---	2230	6
P	2580	2750	5
S	3070	2890	5
Cl	---	5170	3
K	20500	20600	54
Ca	19800	19200	45
Ti	14	37	2
Cr	1.3	(1.5)	0.2
Mn	44	43	2
Fe	528	441	3
Co	0.3	ND	0.4
Ni	1.1	ND	0.4
Cu	8	7.1	0.4
Zn	20	20	0.4
As	0.3	(0.4)	0.1
Se	0.09	ND	0.2
Br	---	3.2	0.1

Rb	---	5.6	0.1
Sr	50	49	0.3
Cd	0.08	ND	1.1
Hg	0.01	ND	0.1
Pb	0.6	ND	0.3

ND = Not Detected () = Below Lower Limit of Quantification

Results: Soils

Soils are mainly inorganic with a low organic component. Sodium, magnesium, aluminum, silicon, potassium, calcium and iron occur at high levels. The results below show soil screening using standardless FP without the use of any reference soils.

Soil Sample 1 Units: ppm (mg/kg) Na, Mg, Al, Si, K, Ca, Fe Units: mass%			
Element	Reference value	NEX CG result	Stat. error
Na	1.060%	0.881%	0.035
Mg	2.148%	2.121%	0.016
Al	6.331%	6.834%	0.008
Si	25.72%	25.33%	0.012
K	1.691%	1.907%	0.009
Ca	6.204%	7.749%	0.014
Fe	3.345%	3.564%	0.003
P	756	852	5
S	195	394	3
Cl	---	89	1
Ti	3839	4200	23
Cr	183	219	4
Mn	734	845	10
Ni	113	128	3
Cu	29	52	1.5
Zn	79	86	1.6
As	9	11	0.6
Se	2.5	2.3	0.3

Br	---	7.6	0.3
Rb	---	112	1
Sr	244	284	1
Cd	0.2	ND	0.9
Sn	3.7	(6.9)	2.0
Ba	342	435	11
Hg	0.04	ND	0.8
Pb	20	26	1


Soil Sample 2			
Units: ppm (mg/kg)			
Na, Mg, Al, Si, K, Ca, Fe Units: mass%			
Element	Reference value	NEX CG result	Stat. error
Na	0.280%	0.237%	0.023
Mg	0.955%	1.018%	0.011
Al	5.455%	6.355%	0.007
Si	19.68%	18.60%	0.010
K	2.298%	2.560%	0.009
Ca	12.89%	15.28%	0.021
Fe	2.499%	2.601%	0.004
P	656	721	4
S	1119	1550	5
Cl	---	102	1
Ti	2771	3220	22
Cr	54	71	3
Mn	322	355	8
Ni	23	34	2
Cu	16	35	1.4
Zn	54	61	1.5
As	19	20	0.7
Se	2.2	2.4	0.3
Br	---	7.2	0.3
Rb	---	129	1

Sr	389	436	2
Cd	0.3	ND	1.0
Sn	3.9	(9.0)	2.0
Ba	288	361	10
Hg	0.02	ND	1
Pb	20	28	1

ND = Not Detected () = Below Lower Limit of Quantification

Fundamental parameters

Rigaku RPF-SQX Fundamental Parameters (FP) and Scattering FP



The Rigaku RPF-SQX software automatically deconvolutes spectral peaks and models the sample matrix and X-ray absorption and enhancement effects using fundamental XRF equations. The versatile RPF-SQX software is simple to use and offers many ways to craft a matrix model based on the specific material composition. This allows for semi-quantitative analysis without the use of any reference standards, typically returning concentration results on the order of 15-20% relative.

Scattering FP is the Rigaku technique of using the measurement of the Compton and Thomson (Rayleigh) scatter peaks to gain valuable information about the sample matrix. By comparing the Compton and Thomson scatter peaks, the average atomic number of the sample is calculated and from this the percentage of the sample that cannot be measured (elements H - F) is estimated, improving the quantification of the elements Na – U. Use of Scattering FP is ideal for the analysis of plant materials and soils where the composition of the unmeasurable balance of the sample can change significantly from sample to sample.

Matching Library

The semi-quantitative measurements using RPF-SQX can be further optimized with the use of a Matching Library, as shown in the results of the plant materials. A Matching Library is easily created by the operator using the measurements of one or more assayed reference samples of the same material type. The measurements of these “type standards” are registered in a library specific for the particular material composition of interest and give the FP theoretical equations

examples of the actual matrix. Depending on the number of type standards and how closely the type standards resemble the material composition of interest, use of a Matching Library can typically improve accuracy to approximately 5-10% relative.

Soils contain much higher levels of inorganic minerals, especially silicates, and therefore can exhibit noticeable mineralogical effects. Soils can vary widely in composition and unmeasurable balance from area to area and region to region. For this reason reference soils to be used as type standards in a Matching Library must come from the same area as the unknown soil samples to be measured in order to ensure optimum accuracy using FP.

The highest degree of accuracy requires the use of matrix-matched empirical regression calibration. For the full characterization of plant materials and soils this may require on the order of a hundred or more calibration standards that model all the variations that may occur, and therefore FP is preferred in the industry rather than empirical calibration.

Conclusion

Given homogenous samples the Rigaku NEX CG yields excellent performance for the elemental analysis of plant materials and soils. If desired, FP standardless semi-quantification can be improved with Matching Libraries based on one or more assayed type standards of the particular material type. The NEX CG software is powerful and flexible, yet simple and intuitive to operate. These features make the NEX CG an ideal EDXRF tool for screening and characterization of plant materials and soils.