View on rigaku.com

XRF1116 - Lead analysis in gasoline — ASTM D5059-21 — using benchtop WDXRF Supermini200

Introduction

Lead (Pb), in the form of alkyl lead such as tetraethyl lead, was used as an antiknock additive to increase the octane rating in gasoline in the past. However, alkyl lead is highly poisonous and gasoline containing alkyl lead, called leaded gasoline, causes air pollution. Nowadays, most countries, except some in Africa, ban leaded gasoline for vehicles. Alkyl lead is still used as an additive in aviation gasoline.

Although lead-free gasoline, called unleaded gasoline, is common in most countries, lead occurs as a contaminant—either unintentionally or intentionally—in some places. Therefore, it is necessary to check the lead concentration in gasoline.

This application note demonstrates quantitative analysis of low concentration lead in gasoline according to ASTM D5059-21 on Rigaku's Supermini200, a benchtop wavelength dispersive X-ray fluorescence (WDXRF) spectrometer.

Instrument

The Supermini200, a tube-below sequential benchtop WDXRF spectrometer, is optimized for the routine analysis that today's petroleum laboratories need to perform. The spectrometer is equipped with a 200 W X-ray tube and analyzing crystals covering 0 to Cm in the standard configuration.

The system software is designed for ease of use in routine analyses. The Flowbar in quantitative analysis guides users in establishing calibration. The Sample ID Table and the Program Operation help operators carry out daily analysis.

Measurements were performed on the Supermini200 at 50 kV and 4 mA using a LiF(200) analyzing crystal.

Analysis method of ASTM D5059-21

The analysis demonstrated in this application note followed Method C of ASTM D5059-21, for low concentration lead, 0.01 g Pb / US gal - 0.5 g Pb / US gal (0.0026 g / L - 0.13 g / L).

In Method C of ASTM D5059-21, the internal standard method with bismuth (Bi) is employed. For each solution of calibration standards or analysis gasoline samples, exactly 20 mL is taken and 2 mL of the Bi internal standard solution (3.00 g Bi / US gal) is added and mixed thoroughly.

X-ray intensities of the following three lines are measured:

• Pb-Lα (at 1.175 Å)

- Bi-Lα (at 1.144 Å)
- Background (1.194 Å)

The intensity ratio R is determined by the following formula:

R = (A - C) / B

```
where A: counting rate at 1.175 Å (Pb-Lα)
B: counting rate at 1.144 Å (Bi-Lα)
C: counting rate at 1.194 Å (background).
```

The corrected ratio RC is determined as follows:

 $R_{C} = R - R_{b}$

where R_b: ratio (R) for blank.

A calibration curve is made with R_C and the slope S of the resulting line is determined as follows:

 $S = (g Pb / US gal) / R_C$.

The lead content of analysis samples is calculated as follows:

Pb content (g / US gal) = $S \times R_C$.

The obtained values of the lead content are reported to the nearest 0.005 g / US gal.

Calibration and standards

"Standards for Lead in Gasoline" (isooctane base), which consists of 0.000 (blank), 0.001, 0.005, 0.010, 0.050, 0.100 and 0.300 g Pb / US gal, and "Internal Standards for XRF Analysis; Bi 0.793 g / L" provided by VHG Labs (LGC Standards) were used to make a calibration curve.

For each calibration standard, exactly 20 mL was taken and 2 mL of the Bi internal standard was added and mixed thoroughly. Approximately 6 g of the mixture of each calibration standard and the Bi internal standard was poured into the liquid cell, Chemplex XRF Sample Cups 1095, with 3.6 µm polyester film.

The X-ray intensities of Pb-La, Bi-La and the background (1.194 Å) were measured for each of the calibration standards. The counting time was 20 seconds for each line. Then, R_C was calculated and the calibration curve was made. The obtained calibration curve was shown in Figure 1.



Figure 1: Calibration curve of lead (Pb) in gasoline by bismuth (Bi) internal standard. The accuracy of calibration is 0.0051 g / US gal.

The accuracy of calibration was calculated by the following formula:

$$Accuracy = \sqrt{rac{\sum_i (C_i - \hat{C}_i)^2}{n-m}}$$

 C_i : certified value of standard sample \hat{C}_i : calculated value of standard sample n : number of standard samples. m: degree of freedom (2: linear)

Repeatability test

For repeatability tests, solutions with approximately 0.01 g Pb / US gal and 0.05 g Pb / US gal were prepared by mixing an organometallic standard of Pb with isooctane.

Repeatability tests were carried out for the above two solutions (~0.01 g Pb / US gal and ~0.05 g Pb / US gal). For the same sample, two aliquots were prepared and quantified with the calibration; this process was repeated twenty times. The test results are tabulated in Table 1, in which the average and the difference of two aliquots each are shown and **r** represents "repeatability" defined in ASTM D5059-21, which states that the difference between successive test results obtained by the same operator with the same apparatus under constant operation conditions on identical test material would, in the long run, in the normal and correct operation of the test method, exceed the following values only in one case in twenty:

r = 0.007 + 0.14 X X: g Pb/US gal.

The test results shown in Table 1, where the difference between two aliquots does not exceed the repeatability (**r**), prove that the performance of the Supermini200 meets the requirement of ASTM D5059-21.

Table 1 Result of repeatability test for ASTM D5059-21.

Sample: ~0.01 g Pb / US gal	(unit: g Pb / US gal)
1 5 5	

Run	1st aliq.	2nd aliq.	Average	Diff.	r (limit)
01	0.015	0.015	0.015	0.000	0.009
02	0.015	0.015	0.015	0.000	0.009
03	0.015	0.015	0.015	0.000	0.009
04	0.015	0.015	0.015	0.000	0.009
05	0.015	0.015	0.015	0.000	0.009
06	0.015	0.010	0.013	0.005	0.009
07	0.010	0.015	0.013	0.005	0.009
08	0.015	0.015	0.015	0.000	0.009
09	0.015	0.015	0.015	0.000	0.009
10	0.015	0.015	0.015	0.000	0.009
11	0.015	0.015	0.015	0.000	0.009
12	0.015	0.015	0.015	0.000	0.009
13	0.015	0.015	0.015	0.000	0.009
14	0.015	0.015	0.015	0.000	0.009
15	0.015	0.015	0.015	0.000	0.009
16	0.015	0.015	0.015	0.000	0.009
17	0.015	0.015	0.015	0.000	0.009
18	0.015	0.015	0.015	0.000	0.009
19	0.015	0.015	0.015	0.000	0.009
20	0.015	0.015	0.015	0.000	0.009

Sample	ample: ~0.05 g Pb / US gal			(unit: g Pb / US gal)	
Run	1st aliq.	2nd aliq.	Average	Diff.	r (limit)
01	0.050	0.050	0.050	0.000	0.014
02	0.050	0.050	0.050	0.000	0.014
03	0.050	0.050	0.050	0.000	0.014
04	0.050	0.055	0.053	0.005	0.014
05	0.055	0.050	0.053	0.005	0.014
06	0.050	0.050	0.050	0.000	0.014
07	0.050	0.050	0.050	0.000	0.014
08	0.050	0.050	0.050	0.000	0.014

090.0500.0500.0500.0000.014100.0500.0500.0500.0000.014110.0500.0500.0500.0000.014120.0500.0500.0500.0000.014130.0500.0500.0500.0000.014140.0500.0500.0500.0000.014150.0500.0500.0500.0000.014160.0500.0500.0500.0000.014170.0500.0500.0500.0000.014180.0500.0500.0500.0000.014190.0500.0550.0530.0050.014						
100.0500.0500.0500.0000.014110.0500.0500.0500.0000.014120.0500.0500.0500.0000.014130.0500.0500.0500.0000.014140.0500.0500.0500.0000.014150.0500.0500.0500.0000.014160.0500.0500.0500.0000.014170.0500.0500.0500.0000.014180.0500.0500.0500.0000.014190.0500.0500.0530.0050.014	09	0.050	0.050	0.050	0.000	0.014
110.0500.0500.0500.0000.014120.0500.0500.0500.0000.014130.0500.0500.0500.0000.014140.0500.0500.0500.0000.014150.0500.0500.0500.0000.014160.0500.0500.0500.0000.014170.0500.0500.0500.0000.014180.0500.0500.0500.0000.014190.0500.0550.0530.0050.014	10	0.050	0.050	0.050	0.000	0.014
120.0500.0500.0500.0000.014130.0500.0500.0500.0000.014140.0500.0500.0500.0000.014150.0500.0500.0500.0000.014160.0500.0500.0500.0000.014170.0500.0500.0500.0000.014180.0500.0500.0500.0000.014190.0500.0500.0530.0050.014	11	0.050	0.050	0.050	0.000	0.014
130.0500.0500.0000.014140.0500.0500.0500.0000.014150.0500.0500.0500.0000.014160.0500.0500.0500.0000.014170.0500.0500.0500.0000.014180.0500.0500.0500.0000.014190.0500.0500.0500.0000.014200.0500.0550.0530.0050.014	12	0.050	0.050	0.050	0.000	0.014
140.0500.0500.0500.0000.014150.0500.0500.0500.0000.014160.0500.0500.0500.0000.014170.0500.0500.0500.0000.014180.0500.0500.0500.0000.014190.0500.0500.0500.0000.014200.0500.0550.0530.0050.014	13	0.050	0.050	0.050	0.000	0.014
150.0500.0500.0500.0000.014160.0500.0500.0500.0000.014170.0500.0500.0500.0000.014180.0500.0500.0500.0000.014190.0500.0500.0500.0000.014200.0500.0550.0530.0050.014	14	0.050	0.050	0.050	0.000	0.014
160.0500.0500.0500.0000.014170.0500.0500.0500.0000.014180.0500.0500.0500.0000.014190.0500.0500.0500.0000.014200.0500.0550.0530.0050.014	15	0.050	0.050	0.050	0.000	0.014
170.0500.0500.0500.0000.014180.0500.0500.0500.0000.014190.0500.0500.0500.0000.014200.0500.0550.0530.0050.014	16	0.050	0.050	0.050	0.000	0.014
18 0.050 0.050 0.000 0.014 19 0.050 0.050 0.000 0.014 20 0.050 0.055 0.053 0.005 0.014	17	0.050	0.050	0.050	0.000	0.014
19 0.050 0.050 0.000 0.014 20 0.050 0.055 0.053 0.005 0.014	18	0.050	0.050	0.050	0.000	0.014
20 0.050 0.055 0.053 0.005 0.014	19	0.050	0.050	0.050	0.000	0.014
	20	0.050	0.055	0.053	0.005	0.014

Conclusion

Lead in gasoline can be routinely analyzed with high accuracy and precision on the Supermini200, a benchtop sequential WDXRF spectrometer. This application note demonstrates that the performance of the Supermini200 meets the requirement of ASTM D5059-21.

Reference

ASTM 5059-21 Standard Test Methods for Lead in Gasoline by X-Ray Spectroscopy

Related products



Supermini200

Benchtop tube below sequential WDXRF spectrometer anal yzes O through U in solids, liquids and powders